



CERTIFICATE OF ACCREDITATION

The ANSI National Accreditation Board

Hereby attests that

LexaMed, Ltd.

**705 Front Street
Toledo, OH 43605**

Fulfills the requirements of

ISO/IEC 17025:2017

In the field of

TESTING

This certificate is valid only when accompanied by a current scope of accreditation document.
The current scope of accreditation can be verified at www.anab.org.

Jason Stine, Vice President

Expiry Date: 01 August 2027

Certificate Number: AT-2689



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

LexaMed, Ltd.

705 Front Street
Toledo, OH 43605

Anne Schuler 419-693-5307
aschuler@lexamed.net www.lexamed.net

Brandi Heckman 419-693-5307
bheckman@lexamed.net www.lexamed.net

TESTING

ISO/IEC 17025 Accreditation Granted: **01 August 2025**

Certificate Number: **AT-2689**

Certificate Expiry Date: **01 August 2027**

Microbiological

Specific Tests and/or Properties Measured	Specification, Standard, Method, or Test Technique	Items, Materials or Product Tested	Key Equipment or Technology
Antimicrobial Preservative Effectiveness	SOP LEXLP-072 based on USP <51>	Antimicrobial Preservatives	ISO Class 5 LAF Hood Incubators
Bacterial Endotoxin (Gel-clot)	SOP LEXLP-004 and LEXLP-003 based on: USP <85>; USP <161>; ANSI/AAMI ST72	Medical Devices, Pharmaceuticals	Depyrogenation oven
Bacterial Endotoxin (Kinetic Turbidimetric)	SOP LEXLP-085 and LEXLP-091 based on: USP <85>; USP <161>; ANSI/AAMI ST72	Medical Devices, Pharmaceuticals	Depyrogenation oven Microplate reader
Bioburden	SOP LEXLP-014 and LEXLP-037 based on: ANSI/AAMI/ISO 11737-1; USP <61>; USP <1111>	Medical Devices, Pharmaceuticals	ISO Class 5 LAF Hood Incubators
Cytotoxicity (MEM)	SOP LEXLP-047 based on: USP <87>; ISO 10993-5	Medical Devices	ISO Class 5 LAF Hood Microscope Incubators
Cytotoxicity Agar Overlay)	LEXLP-048 based on: USP <87>; ISO 10993-5	Medical Devices	ISO Class 5 LAF Hood Microscope Incubators

This Scope of Accreditation, version 006, was last updated on: 31 July 2025 and is valid only when accompanied by the Certificate.


Page 1 of 2

Microbiological

Specific Tests and/or Properties Measured	Specification, Standard, Method, or Test Technique	Items, Materials or Product Tested	Key Equipment or Technology
Particulates	SOP LEXLP-092 and LEXLP-026 based on: USP <788>; USP <789>; USP <1788>	Medical Devices, Pharmaceuticals (ophthalmic solutions/injectables)	Liquid Particle Counter Microscope
Microbiological Examination of Non-Sterile Products – Microbial Limits (Enumeration and Specified Organisms)	LEXLP-038 based on: USP <61>; USP <62>	Pharmaceuticals	ISO Class 5 LAF Hood Incubators
Biological Indicator Population Verification	LEXLP-101 based on: ISO 1138-1	Biological Indicators	ISO Class 5 LAF Hood Incubators
Sterility and B/F Testing	SOP LEXLP-017 and SOP LEXLP-021 based on: USP <71>; ISO 11737-2	Medical Devices, Pharmaceuticals, BIs	ISO Class 5 Cleanrooms and LAF Hoods Incubators

Chemical

Specific Tests and/or Properties Measured	Specification, Standard, Method, or Test Technique	Items, Materials or Product Tested	Key Equipment or Technology
Container Closure Integrity (CCI) Testing (Dye Immersion)	SOP LEXLP-073 based on: USP <1207>	Pharmaceutical Containers	Pressure/Vacuum Vessel, UV/Vis Spectrophotometer
Ethylene Oxide (EO) Residual Analysis	SOP LEXLP-039, LEXLP-040 and SOP LEXLP-070 based on: ISO 10993-7	Medical Devices	Gas Chromatograph (GC)
Protein	SOP LEXLP-033	Medical Devices	UV/Vis Spectrophotometer
Hemoglobin	SOP LEXLP-081	Medical Devices	UV/Vis Spectrophotometer
Carbohydrate	SOP LEXLP-084	Medical Devices	UV/Vis Spectrophotometer



Jason Stine, Vice President